

FIELD OF CALIBRATION : Dimensional Measurement

INSTRUMENTS CALIBRATED	LABORATORY LOCATION		SITE CALIBRATION	
	RANGE	UNCERTAINTY	RANGE	UNCERTAINTY
<b>CALLIPER</b>	0mm - 600mm ( INTERNAL & EXTERNAL)	0.01 mm	0mm - 300mm 301mm - 600mm (INTERNAL ) 0mm - 300mm 301mm - 600mm (EXTERNAL)	0.01mm 0.02mm 0.01mm 0.02mm
<b>MICROMETER</b>	0mm - 25mm	0.001 mm	0mm - 25mm	0.001 mm
<b>HEIGHT GAUGE</b>	0mm - 600mm	0.01 mm	0mm - 300mm 301mm - 600mm	0.01mm 0.02mm
<b>DIAL GAUGE</b>	0mm - 15mm	0.004 mm	0mm - 15mm	0.004 mm
<b>DIGITAL INDICATOR</b>	0mm - 15mm	0.004 mm	0mm - 15mm	0.004 mm
<b>PROFILE PROJECTOR</b>			0mm - 50mm	0.006 mm

FIELD OF CALIBRATION : Torque

INSTRUMENTS CALIBRATED	LABORATORY LOCATION		SITE CALIBRATION	
	RANGE	UNCERTAINTY	RANGE	UNCERTAINTY
<b>TORQUE SCREWDRIVERS / WRENCHES</b>	15kgf.cm - 115kgf.cm	0.9kgf.cm		